

# TransFix<sup>®</sup>

## Certificate of Analysis

### TransFix<sup>®</sup>/EDTA Vacuum Blood Collection Tubes (TVT)

**RUO** **REF** : TVT-09-1-RUO, TVT-09-2-RUO, TVT-09-50-RUO **LOT** : 30054

 : 2025-12-31

Start Date of Manufacture: **2024-05-30**

TransFix Batch Number used in TVT manufacture: **30047**

Start Date of Quality Control Testing: **2024-06-20**

	A	B		
	Control Day 0	TVT Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	83.4	83.1	0.3	PASS
% CD4+	49.8	51.3	3.1	PASS
CD3+ Abs Cnt	2183	2080	4.7	PASS
CD4+ Abs Cnt	1087	1070	1.6	PASS
CD8+ Abs Cnt	868	901	3.8	PASS
CD45+ Abs Cnt	2662	2560	3.8	PASS

#### Pass Criterion:

The pass criterion is a less than 15% difference between the blood lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

#### Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 04 Jul 2024