

# TransFix<sup>®</sup>

## Certificate of Analysis TransFix<sup>®</sup> Cellular Antigen Stabilisation Reagent

**LOT** : 30060

: 2026-08-31

Description: **TransFix Bulk**Reference: **TFB-01-1, TFB-01-10, TFB-01-50 and TFB-20-1**Start Date of Manufacture: **2024-08-08**Start Date of Quality Control Testing: **2024-09-05**

	A	B		
	Control Day 0	TransFix Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	78.7	79.1	0.5	Pass
% CD4+	64.7	64.5	0.3	Pass
CD3+ Abs Cnt	1438	1407	2.1	Pass
CD4+ Abs Cnt	930	908	2.4	Pass
CD8+ Abs Cnt	498	478	3.9	Pass
CD45+ Abs Cnt	1862	1810	2.8	Pass

**Pass Criterion:**

The pass criterion is a less than 15% difference between the lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

**Comments:**

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 18 Sep 2024